
	Record	N°	520-FOR
		Revision	1
		Page	1/7
SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test			

1. Introduction

The Summary of Safety and Performance (SSP) is one of the requirements of the new Regulation (IVDR 2017/746), specific for class C and D devices, to enhance transparency and adequate access to information. It intends to provide public access to summarised data on the safety and performance of class C and class D IVD devices to all intended users – professionals and lay persons. GaDia is also providing this information for class B devices.

2. Summary of Safety and Performance (SSP)


Requirements based on IVDR Article 29	Potential regulatory sources
Device identification and general information	
Name or trade name including any model number or version	KarbaDia Rapid Test
Manufacturer (name and address)	GaDia SA Route de l'Île-au-Bois 1A 1870 Monthey Switzerland
Manufacturers single registration number (SRN), if available	CH-MF-000031123
Basic UDI-DI	7649990065KARMM
Intended purpose of the device	
Intended purpose and indications	KarbaDiag is a non-automated rapid immunochromatographic test intended to be used for the qualitative detection of KPC-type, NDM-type, IMP-type, VIM-type and OXA-48-type carbapenemase in bacterial colonies. The assay is for professional use only and can aid in the diagnosis of KPC-type, NDM-type, IMP-type, VIM-type and OXA-48-type carbapenemase resistant strains. The test should be used in conjunction with other diagnostic procedures, such as genetic analysis, susceptibility testing and other microbial analysis.
Target populations	Bacterial colonies on agar plate suspected to be Carbapenem resistance Enterobacteriaceae. Bacteria can be isolated from human body fluids or other sources.

	Record	N°	520-FOR
		Revision	1
		Page	2/7
SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test			

<p>Contraindications (limitations)</p>	<ol style="list-style-type: none"> 1. The test procedure, precautions and interpretation of results for this test must be followed strictly when testing. 2. The test is for professional use only and not for home-testing or near- patient testing. 3. The product is only used for the detection of Candida antibody in serum, plasma and BAL samples. 4. The test results of this kit are for reference only and should not be used as the only basis for clinical diagnosis and treatment. The clinical management of patients should be comprehensively considered in conjunction with their symptoms, medical history, other laboratory tests and treatment responses. 5. The sample treatment solution must be clear and without turbidity.
--	--

Device description

<p>Device description</p>	<p>KarbaDia is a sandwich immunochromatographic assay. The test has 5 pre-coated test lines (K N I V O) on nitrocellulose membrane and one Control line (C) per test. If KPC-type, NDM-type, IMP-type, VIM-type or OXA-48-type carbapenemase are present in the specimen, it will bind to the gold-conjugated anti-KPC, anti-NDM, anti-IMP, anti-VIM or anti-OXA-48 antibodies, respectively, pre-dried on conjugate pad. The gold-conjugated antibody-antigen complex moves upward on the membrane by capillary action where it will react with the test lines. The immobilized anti-KPC, anti-NDM, anti-IMP, anti-VIM or anti-OXA-48 monoclonal antibodies on test lines will capture the gold-conjugated antibody-antigen complex and form red line(s). Whether the sample is positive or not, the nanoparticle complex continues to move across the membrane where immobilized goat anti-chicken IgY antibodies (control line) will bind the gold-conjugated chicken IgY antibodies and form a visible red control line. Positive test results form one</p>
---------------------------	---

	Record	N°	520-FOR
		Revision	1
		Page	3/7
SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test			

	or more red line(s) in test area. Negative test results form only a control line (C line). The quality control line (C line) is an internal quality control that serve as (1) verification that sufficient volume is added, (2) that proper flow is obtained and (3) an internal control for the reagents. The control line must always appear.
Reference to previous generation(s) or variants of the device (as applicable) and a description of the differences	N/A
Description of accessories intended to be used in combination with the device (as applicable)	N/A
Description of other devices and products intended to be used in combination with the device (as applicable)	Materials Required but Not Supplied 1. Timer 2. Inoculation loop (5 µl) 3. Optional: Pipettes and sterile tips 4. Optional: Disposable sterile micro-centrifuge tubes (1.5 ml)
Standards Reference	
Harmonised standards and Common Specifications (CS) applied	IVDD 98/79/EC EN ISO 13485:2016 EN ISO 15223-1:2021 EN ISO 17511:2021 ISO 14971:2019 ISO 18113-1:2009 ISO 18113-2:2009 ISO 20417:2021 ISO 13975:2003 ISO 13612:2002 ISO 23640:2011 ISO 20916:2019 IEC 62366-1:2015+A1:2020
Summary of the Performance Evaluation	
A total of 212 clinical isolates were collected from European hospitals (81%) and American hospitals (15%) in 2019 and 2020, including 19 carbapenemase- negative isolates (9%) and 193 (91%) carbapenem-resistant Enterobacterales (CRE) with various carbapenemase types. These were collected from various infection sources including respiratory, urinary tract, intra-abdominal and chorionic villus sampling. Molecular analysis was performed on all isolates and MIC determination on discordant results. The isolates were cultivated on	

SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test

blood agar for 24h at 37°C and analysed with KarbaDia rapid test. The primary end point was to assess the diagnostic performance of KarbaDia Rapid test (GaDia SA, Monthey, Switzerland) using cultivated bacterial colonies. The second end point was to assess the capacity of the test to detect specific variants and non-fermenter isolates. The usability of the test and feedbacks from user are also collected for the usability study. Vassarstats online tool (www.vassarstats.net) was used to calculate sensitivity (SE), specificity (SP), positive and negative predictive values (PPV, NPV), 95% confidence intervals, median, and Interquartile range (IQR); while significance (p-values) was calculated using student t test for independent samples with equal variances. Statistical significance was defined as $p < 0.05$.

Total Isolates	203		
Carbapenemase negative isolates	19		
KPC isolates	32		
KPC-2	20 (63%)		
KPC-3	8 (25%)		
KPC-4	3 (9%)		
KPC-27	1 (3%)		
OXA isolates	61		
OXA-48-TYPE(u)	2 (3%)		
OXA-48(c)	31 (51%)		
OXA-181(c)	11 (18%)		
OXA-232(c)	16 (26%)		
OXA-244(c)	1 (2%)		
NDM isolates	69		
NDM-1	49 (71%)		
NDM-2	7 (10%)		
NDM-5	10 (14%)		
NDM-6	1 (1%)		
NDM-7	1 (1%)		
NDM-Type	1 (1%)		
IMP isolates	13		
IMP-1	3 (23%)		
IMP-4	2 (15%)		
IMP-8	4 (31%)		
IMP-26	1 (8%)		
IMP Multicopy	2 (15%)		
IMP-Type	1 (8%)		
VIM isolates	29		
VIM-1	24 (83%)		
VIM-2	4 (14%)		
VIM-Type	1 (3%)		
Carbapenemase negative (n=19)		Carbapenemase positive (n=184)	
Europe	12 (63%)	Europe	149 (81%)
America	7 (37%)	America	27 (15%)
Asia	0 (0%)	Asia	6 (3%)
Pacific	0 (0%)	Pacific	2 (1%)
Enterobacteriaceae	19 (100%)	Enterobacteriaceae	139 (76%)
Age (average)	64	Age (average)	56
Female	8 (42%)	Female	62 (34%)
CVS isolation	5 (26%)	CVS isolation	45 (24%)
UTI isolation	2 (11%)	UTI isolation	35 (19%)
RTI isolation	6 (32%)	RTI isolation	69 (38%)
IAI isolation	6 (32%)	IAI isolation	34 (18%)

SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test

KPC	+	-	Sensitivity 100% (CI95%: 87-100%)	
	32	0		Specificity 100% (CI95%: 97-100%)
	0	180		
			PPV 100% (CI95%: 87-100%)	
			NPV 100% (CI95%: 97-100%)	
OXA	+	-	Sensitivity 98% (CI95%: 90-100%)	
	58	2		Specificity 99% (CI95%: 95-100%)
	1	149		
			PPV 97% (CI95%: 87-99%)	
			NPV 99% (CI95%: 96-100%)	
NDM	+	-	Sensitivity 97% (CI95%: 89-99%)	
	65	1		Specificity 99% (CI95%: 95-100%)
	2	139		
			PPV 98% (CI95%: 91-100%)	
			NPV 99% (CI95%: 94-100%)	
IMP	+	-	Sensitivity 93% (CI95%: 66-100%)	
	14	0		Specificity 100% (CI95%: 98-100%)
	1	190		
			PPV 100% (CI95%: 73-100%)	
			NPV 99% (CI95%: 97-100%)	
VIM	+	-	Sensitivity 100% (CI95%: 85-100%)	
	29	0		Specificity 100% (CI95%: 97-100%)
	0	183		
			PPV 100% (CI95%: 85-100%)	
			NPV 100% (CI95%: 97-100%)	

Carbapenemase Type	Variants detected (previous study)	Variants detected (this study)
KPC	KPC-1, KPC-2, KPC-3, KPC-74	KPC-2, KPC-3, KPC-4, KPC-27
OXA	OXA-23 OXA-163, OXA-181, OXA-232	OXA-48(c), OXA-48 Type(u) OXA-181(c), OXA-232(c) OXA-244(c)
NDM	NDM-1, NDM-5, NDM-7	NDM-1, NDM-2, NDM-5, NDM-6, NDM-7
IMP	IMP-1, IMP-3, IMP-4, IMP-6, IMP-10, IMP-25, IMP-26, IMP-30, IMP-34, IMP-38, IMP-40, IMP-42	IMP-1, IMP-4, IMP-6, IMP-7, IMP-10, IMP-26
VIM	VIM-1, VIM-2, VIM-4, VIM-5, VIM-9, VIM-10	VIM-1, VIM-2

The main finding of this evaluation study, using an unmatched case control design including 91% (184/203) carbapenemase positive isolates is that the diagnostic performance of KarbaDia RDT on bacterial isolates is high and depend of the carbapenemase type detected. The sensitivity varied between 93% and 100% with lower sensitivity with IMP type, when comparing with genetic and MIC assay. In term of specificity, the diagnostic specificity is between 99% and 100%, similar to other rapid tests.


In conclusion, KarbaDia gives valuable information in only 15 minutes to start the treatment earlier with the right antibiotic. Performance compared to genetic and MIC methods, longer and requiring laboratory equipment, has substantial agreement. Confirmation with laboratory methods should be conducted.

Summary of the Post-Market Performance Follow-Up

Two published studies use the same device and bring important information about performance of the test. Findings of these 2 studies are summarized below.

Sadek et al. Diagnostic Microbiology and Infectious diseases

The Rapid carbapenemase test is a novel immunochromatographic test for detection of the 5 major carbapenemases (KPC, NDM, IMP, VIM, and OXA-48-like). This test is rapid, easy

	Record	N°	520-FOR
		Revision	1
		Page	6/7
SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test			

to perform, and shows a good sensitivity and specificity (96.8% and 100%, respectively), being suitable for microbiology laboratories together with biochemical rapid tests.

The study tested different variants of carbapenemase that need to be included in the table in IFU and in Performance evaluation report. The following table indicates the additional variants detected in the study of Sadek et al. 2022)

Carbapenemase type	Variant detected (Sadek et al)
KPC	KPC-2, KPC-3
NDM	NDM-1, NDM-5, NDM-6, NDM-7, NDM-9, NDM-24
VIM	VIM-1, VIM-2, VIM-4, VIM-19, VIM-53
IMP	IMP-1, IMP-4, IMP-5, IMP-10, IMP-11, IMP-15, IMP-29
OXA	OXA-48, OXA-162, OXA-181, OXA-204, OXA-244

Zhang et al. antibiotics MDPI

Rapid and accurate detection can help optimize patient treatment and improve infection control against nosocomial carbapenemase-producing organisms (CPO). In this study, a total of 217 routine clinical isolates (Enterobacterales and *A. baumannii*), including 178 CPOs and 39 non- CPOs, were tested to evaluate the performance of six phenotypic carbapenemase detection and classification assays, i.e., BD Phoenix CPO detect panel, Rapidec Carba-NP, O.K.N detection kit, and three carbapenem inactivation methods (CIMs; mCIM, eCIM, sCIM).

Table 2. Overall sensitivity and specificity of phenotype diagnostic assays.


Diagnostic Assays		Sensitivity		Specificity	
		%	95% CI	%	95% CI
BD Phoenix CPO detect panel	P/N test	98.78	95.21–99.79	79.49	63.06–90.13
	Ambler test	56.71	48.75–64.34	94.87	81.37–99.11
	Rapidec Carba-NP	91.93	86.30–95.45	100	88.83–100
CPO detection tests	mCIM	98.06	94.00–99.50	97.44	84.92–99.87
	sCIM	96.89	92.52–98.85	94.87	81.37–99.11
	O.K.N Detection kit	99.28	95.43–99.96	100	88.83–100
CPO classification tests	mCIM + eCIM	92.90	87.35–96.23	97.44	84.92–99.87

Table 4. Performance of other carbapenemase detection and classification diagnostic assays.

Ambler Class	Carbapenemase	CPO Detection Tests			CPO Classification Tests	
		Rapidec Carba-NP	mCIM	sCIM	O.K.N Detection Kit	mCIM + eCIM
ClassA						
ClassB	KPC (n = 59)	59	59	58	59	55
	NDM (n = 52)	52	52	52	51	52
	IMP (n = 15)	15	15	15	0	12
	VIM (n = 1)	1	1	1	0	1
ClassD	SIM (n = 1)	1	1	1	0	1
	OXA-23 (n = 5)	5	/	1	0	/
	OXA-58 (n = 1)	1	/	1	0	/
Dual enzymes	OXA-48-like (n = 16)	5	13	16	16	12
	Class A + B					
	KPC + NDM (n = 4)	4	4	4	4	4
Class B + B non-CPO	KPC + IMP (n = 2)	0	2	2	2	2
	NDM + IMP (n = 5)	5	5	5	5	5
	(n = 39)	0	1	2	0	1

In this study, OXA-23 and OXA-58 were not detected as well as 1 strain with NDM variant (unknown variant).

Metrological traceability	
Metrological traceability of assigned values	N/A

	Record	N°	520-FOR
		Revision	1
		Page	7/7
SUMMARY OF SAFETY AND PERFORMANCE (SSP) KarbaDia Rapid Test			

Users	
User Profile	The tests can be performed in laboratories by health care workers or laboratory technicians with appropriate training in sample collection, biosafety and in the use of rapid tests.
User Training	Appropriate training in sample collection, biosafety and in the use of rapid tests.
Device Risks Information	
Residual risks and undesirable effects	<ul style="list-style-type: none"> - Contamination of the user by infected samples - Wrong interpretation of the test results - False negative
Warnings and precautions	<ol style="list-style-type: none"> 1. Read the instruction for use carefully before using the test. 2. Clearly identify the sample ID on the test cassettes. 3. This product is for in vitro diagnostic and professional use only. 4. Do not reuse the test 5. Do not use the test after expiry date 6. Read the test results within the specific time to avoid wrong interpretation. 7. Do not use the components from different batches or different types of reagents. 8. Properly dispose the specimen and used materials following the local biohazardous disposal regulation. 9. Use protective equipment when using the test and handling samples as they may contain infectious agents, human or animal components. 10. Sodium azide is used as preservative in the sample treatment solution. Dispose material according to relevant local regulations and avoid contact with eyes and skin. 11. Any serious incident that has occurred in relation to the device shall be reported to the manufacturer and the competent authority of the Member State in which the user and/or the patient is established.